

Volumetric flasks, boro 3.3, class A, amber graduations, with PP stoppers, incl. ISC

All Blaubrand[®]: volumetric flasks are delivered with batch number and one batch certificate included per pack. They are also available with DAkkS calibration certificate upon request.

- BLAUBRAND[®]: ETERNA
- DIN EN ISO 1042
- DE-M marking
- Graduation and inscriptions in Eterna amber stain
- Calibrated to contain (IN)

Standard: DIN EN ISO 1042

Nominal capacity ml	Tolerance ± ml	Ground size	Accuracy class	Material	Material closure	Type of closure	PK	Cat. No.
5	0,04	NS10/19	A	Boro 3.3	PP	Stopper	1	4.684 457
10	0,04	NS10/19	A	Boro 3.3	PP	Stopper	1	4.684 458
20	0,04	NS10/19	A	Boro 3.3	PP	Stopper	1	4.684 459
20	0,06	NS12/21	A	Boro 3.3	PP	Stopper	1	4.684 460
25	0,04	NS10/19	A	Boro 3.3	PP	Stopper	1	4.684 461
25	0,06	NS12/21	A	Boro 3.3	PP	Stopper	1	4.684 462
50	0,06	NS12/21	A	Boro 3.3	PP	Stopper	1	4.684 463
50	0,10	NS14/23	A	Boro 3.3	PP	Stopper	1	6.223 199
100	0,10	NS14/23	A	Boro 3.3	PP	Stopper	1	4.684 464
200	0,15	NS14/23	A	Boro 3.3	PP	Stopper	1	4.684 465
250	0,15	NS14/23	A	Boro 3.3	PP	Stopper	1	4.684 466
500	0,25	NS19/26	A	Boro 3.3	PP	Stopper	1	4.684 467
1000	0,40	NS24/29	A	Boro 3.3	PP	Stopper	1	4.684 468
2000	0,60	NS29/32	A	Boro 3.3	PP	Stopper	1	4.684 469

Alternativen:

Volumetric flasks, borosilicate glass 3.3, Class B. / BRAND

Volumetric flasks, boro 3.3, class A, blue graduations / BRAND

Volumetric flasks for sugar tests, Borosilicate glass 3.3, class B, white graduated / BRAND

Volumetric flasks, boro 3.3, class A, amber graduations, with PP stopper, with DAkkS calibration certificate / BRAND

Volumetric flasks, boro 3.3, class A, amber graduations, with PP stopper / BRAND